

Reference your bid dated 26/10/2017 on subject cited above.

It is to inform you that the process of technical evaluation has been completed and this draft Technical Evaluation Report is hereby shared/ announced as a pre-requisite of the procurement process.

The attached proformas indicating marks given to your bid/s is sent to you for your information, clarification and proposing corrections, if any.

This draft report is open to rectification based on typographical/ computational errors and / or omissions & merit based complaint, if any, in order to provide equal opportunity, ensure transparency, fairness and provision of maximum playing field to all bidders for a healthy competition in the public interest.

Please note that during the course of evaluation of your bid/s, certain information was not readily appreciable and thus zero marks were awarded provisionally which is subject to the clearance by T&E committee.

Further to note that this report relates to the items pertaining to category A, B ,E & F and the results of remaining categories i.e. C & D will be announced once their evaluation is completed. Bidders who have applied for leftover categories will be informed accordingly.

You are, therefore, supposed to propose corrections for errors/ omissions in technical evaluation report on our part and not to propose any alterations in the bids already submitted by your firm till 14/12/2017 during office hours.

**Evolution Criteria for the Purchase of Category A (Left over Items) (Section-IV of Statement of Requirement) for the financial year 2017-18**

| Sl# | No of Firms | Firm Name               | 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9 | 10 | 11 | 12 | 13 | 14 | 15                                  | 16 | 17 | 18 | 19 | 20 | 21 | 22  | 23 | 24 | 25             |  |   |   |  |  |  |  |  |   |   |  |   |                                     |   |  |  |   |   |   |
|-----|-------------|-------------------------|---|---|---|---|---|---|---|---|---|----|----|----|----|----|-------------------------------------|----|----|----|----|----|----|-----|----|----|----------------|--|---|---|--|--|--|--|--|---|---|--|---|-------------------------------------|---|--|--|---|---|---|
|     |             |                         |   |   |   |   |   |   |   |   |   |    |    |    |    |    |                                     |    |    |    |    |    |    |     |    |    |                | Firm Evaluation Parameters   |   |   |  |  |  |  |  |   |   |  |   | Product Evaluation Parameter        |   |  |  |   |   | Technical Marks   |
|     |             |                         |   |   |   |   |   |   |   |   |   |    |    |    |    |    |                                     |    |    |    |    |    |    |     |    |    |                | Personnel/ Human Resource  |   |   | Workshop facility Testing/Calibration of Equipment |  |  | Manufacturer Performance   |  |   | Conformance To Specification                  |  |   | Product International Certification |   |  | Product Local Performance                      |   |   | Technical Marks out of 100  |
|     |             |                         |   |   |   |   |   |   |   |   |   |    |    |    |    |    |                                     |    |    |    |    |    |    |     |    |    |                | Simple Technician (6/12 months certificate)  | Associate Engineer with Diploma of Associate Engineer | Graduate Engineer duly registered with PEC/MSc Electronics with 5years experience | Availability of workshop in khyber Pakhtunkhwa     | Availability of workshop at National level | List of related tools for the quoted items on firm's pad | Detail of Spare parts availability on the firm's pad(for the quoted items) | Availability of testing tools for calibration by the firm (for quoted items) | Valid ISO 14003 (Environment Certificate) | Valid ISO 9001 Quality Management Certificate | Valid ISO 13485 Quality management certificate | Fully compliance with the required specifications as per statement of Requirement (Up to a maximum of four Minor deviations may be accommodated subject to the condition that main function and performance in any aspect would not affect. However, up to four marks will be deducted) | Additional features of the product  | Certificate of US Food and Drug Administration (US FDA) | Certificate of European Community MDD (CE MDD) | Certificate of Japan Industrial Standard (JIS) | Weightage for local Pakistani original manufacturer | One mark for each satisfactory performance certificate for quoted model or at least the previous provided model of equipment from the public sector medical institution of Pakistan | One mark for each satisfactory performance certificate for quoted model or at least the previous provided model of equipment from the teaching level private sector medical institution of Pakistan |
| --- | ---         | ---                     | 2 | 2 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 2  | 2  | 2  | 40 | 6  | No marks, being mandatory parameter |    |    | 4  | 3  | 3  | 4  | 100 |    |    |                |  |   |   |  |  |  |  |  |   |   |  |   |                                     |   |  |  |   |   |   |
| 1   | 1           | LATIF BROTHERS          | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0  | 0  | 0  | 0  | 0                                   | 0  | 0  | 0  | 0  | 0  | 0  | 0   | 0  | 0  | Non Responsive | Quoted Alternate model   |   |   |  |  |  |  |  |   |   |  |   |                                     |   |  |  |   |   |   |
|     | 2           | Radiant Medical Pvt Ltd | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0  | 0  | 0  | 0  | 0                                   | 0  | 0  | 0  | 0  | 0  | 0  | 0   | 0  | 0  | Non Responsive | Authorization letter is not Embassy Attested.  |   |   |  |  |  |  |  |   |   |  |   |                                     |   |  |  |   |   |   |
| 2   | 1           | MEDI LINE TECHNOLOGY    | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0  | 0  | 0  | 0  | 0                                   | 0  | 0  | 0  | 0  | 0  | 0  | 0   | 0  | 0  | Non Responsive | Authorization is not attested by embassy. Not according to specifications quoted pulse width frequency is 0.5-0.75ms while required is 1.5ms. Stimulus current is fixed to 800mA, while required is 500 to 800 mA. Charge greater than 5mC on one side |   |   |  |  |  |  |  |   |   |  |   |                                     |   |  |  |   |   |   |
|     | 1           | IBS Pharmaceutical      | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0  | 0  | 0  | 0  | 0                                   | 0  | 0  | 0  | 0  | 0  | 0  | 0   | 0  | 0  | Non Responsive | Authorization is not from MF Equipmentos (Manufacturer). Date of Embassy attestation of authorization is after closing date for bid i.e 26.10.2017   |   |   |  |  |  |  |  |   |   |  |   |                                     |   |  |  |   |   |   |



Evaluation Criteria for the Purchase of Category-B (Left over items) for the Financial Year 2017 -18

| Sl | No of Firms Submitted bid | 1                  | 2  | 3                          |   |   |  |  |  |  |   |   |   | 14   |                                    |   |   |  |   |   |  |   |                       | 23 | 24 | 25  |                            |  |  |
|----|---------------------------|--------------------|--|----------------------------|---|---|--|--|--|--|---|---|---|--|------------------------------------|---|---|--|---|---|--|---|-----------------------|----|----|-----|----------------------------|--|--|
|    |                           |                    |  | Firm Evaluation Parameters |   |   |  |  |  |  |   |   |   | Product Evaluation Parameter                   |                                    |   |   |  |   |   |  |   |                       |    |    |     | Technical Marks            |  |  |
|    |                           |                    |  | Personnel/ Human Resource  |   | Workshop facility Testing/Calibration of Equipment  |  |  |  | Manufacturer Performance   |   |   |   | Conformance To Specification                   |                                    | Product International Certification                     |   |  |   | Product Local Performance   |  |   |                       |    |    |     | Technical Marks out of 100 |  |  |
|    |                           |                    |  | Simple Technician          | Associate Engineer with Diploma of Associate Engineer | Graduate Engineer duly registered with PEC/NASC-Electronics with minimum five years' experience | Availability of fully functional works hop in Khyber Pakhtunkhwa | Availability of fully functional works hop at National level | List of related tools for the quoted items on firm's pad | Detail of Spare parts availability on the firm's pad(for the quoted items) | Deduction/certificate of calibration by the firm (for quoted items) | Valid ISO 14001 Environment Certificate | Valid ISO 9001 Quality Management Certificate | Valid ISO 13485 Quality management certificate | Additional features of the product | Certificate of US Food and Drug Administration (US FDA) | Certificate of European Community MDD (CEMDD) | Certificate of Japan Industrial Standard (JIS) | Weightage for local Pakistani original manufacturer | One mark for each satisfactory performance certificate for quoted model or at least the previous provided model of equipment from the public sector medical institution of Pakistan | One mark for each satisfactory performance certificate for quoted model or at least the previous provided model of equipment from the private sector medical institution of Pakistan | Warranty Period three years both with spare parts and services & next two year services without parts from the date of finalisation | Total Technical Score |    |    |     | Status                     | Remarks  |  |
|    |                           |                    |  | 2                          | 2   | 5   | 5  | 5  | 5  | 5  | 5   | 5                                       | 2   | 2  | 2                                  | 30  | 4   | 4  | 4   | 4   | 4  | 4   | 3                     | 3  | 4  | 100 |                            |  |  |
| 1  | 1                         | Paradise Export Co | Air Drill With Cylinder/ Orthopedic Pneumatic Drill with Cylinder Dimend Germany Air driven Handpiece Number 33.326.05   | 0                          | 0   | 0   | 0  | 0  | 0  | 0  | 0   | 0                                       | 0   | 0  | 0                                  | 0   | 0   | 0  | 0   | 0   | 0  | 0   | 0                     | 0  | 0  | 0   | Non responsive             | Not According to specification. Quoted speed is less then the required 900-1250rpm, Cylinder not quoted. |  |
| 2  | 1                         | JATIF BROTHERS     | APPLANATION TONOMETER MODEL :AT-1 MAKE: TAKAGI JAPAN   | 2                          | 2   | 5   | 5  | 5  | 5  | 5  | 5   | 0                                       | 0   | 0  | 30                                 | 0   | 0   | 0  | 0   | 0   | 0  | 0   | 0                     | 0  | 4  | 68  | Responsive                 | NIL  |  |
|    | 2                         | IBS Pharmaceutical | Applanation Tonometer, Model:Y230R, Make:ARI China.  | 2                          | 2   | 5   | 5  | 5  | 5  | 5  | 5   | 0                                       | 0   | 0  | 30                                 | 0   | 0   | 0  | 0   | 0   | 0  | 0   | 0                     | 0  | 4  | 68  | Responsive                 | Nill   |  |
|    | 3                         | SUDAIS ASSOCIATES  | APPLANATION TONOMETER MODEL: AT-R MAKE RAY VISION CHINA  | 2                          | 2   | 5   | 5  | 5  | 5  | 5  | 5   | 0                                       | 0   | 0  | 30                                 | 0   | 0   | 0  | 0   | 0   | 0  | 0   | 0                     | 0  | 4  | 68  | Responsive                 | NIL  |  |
| 3  | 1                         | Medco              | Autodave Automatic 100 L Model: BKQ-8Z 100(H) Make Biobase China   | 0                          | 0   | 0   | 0  | 0  | 0  | 0  | 0   | 0                                       | 0   | 0  | 0                                  | 0   | 0   | 0  | 0   | 0   | 0  | 0   | 0                     | 0  | 0  | 0   | Non Responsive             | Not according to specification, Pressure is less then Required. Quoted Manual doors                      |  |
|    | 2                         | Clinical Life Inc  | Autodave Automatic 100 Liters Model Name: Steri 21 Model number: 40.40.66 Manufacturer: A.J.Costa Origen: Portugal Ro system Model: DM-50-2A   | 2                          | 2   | 5   | 0  | 5  | 5  | 5  | 5   | 0                                       | 2   | 2  | 30                                 | 0   | 0   | 4  | 0   | 0   | 0  | 0   | 0                     | 0  | 4  | 71  | Responsive                 | NIL  |  |
|    | 3                         | Delta Plus Pvt Ltd | Autoclave Automatic 100 Liters Model: Azteca AC-470 Manufacturer: Celttron Medical Technology kft Origen:p Hungry (E.U) RO System: FIZICO 100L Manufacturer: FIZICO-Z.A Engineers Lahore (Dow USA parts) | 0                          | 0   | 0   | 0  | 0  | 0  | 0  | 0   | 0                                       | 0   | 0  | 0                                  | 0   | 0   | 0  | 0   | 0   | 0  | 0   | 0                     | 0  | 0  | 0   | Non Responsive             | Authorization letter is not Embassy Attested.  |  |
|    | 4                         | RAY TECHNOLOGIES   | AUTOCLAVE AUTOMATIC 100 LITERS MODEL:113 HIRAYAMA MAKE:JAPAN   | 0                          | 0   | 0   | 0  | 0  | 0  | 0  | 0   | 0                                       | 0   | 0  | 0                                  | 0   | 0   | 0  | 0   | 0   | 0  | 0   | 0                     | 0  | 0  | 0   | Non Responsive             | Not according to specifications. Separate Steam Generator Quoted   |  |
|    | 5                         | Total Technology   | Autodave Automatic 100 Liters Model: Clinclav 4S Make: Melag Gemany  | 0                          | 0   | 0   | 0  | 0  | 0  | 0  | 0   | 0                                       | 0   | 0  | 0                                  | 0   | 0   | 0  | 0   | 0   | 0  | 0   | 0                     | 0  | 0  | 0   | Non Responsive             | Not according to specification, Pressure is less then required.  |  |
|    | 6                         | KASBN              | AUTOCLAVE AUTOMATIC 100 LTR MODEL: AS-446 WPA MAKE: SMSM POLAND  | 0                          | 0   | 0   | 0  | 0  | 0  | 0  | 0   | 0                                       | 0   | 0  | 0                                  | 0   | 0   | 0  | 0   | 0   | 0  | 0   | 0                     | 0  | 0  | 0   | Non Responsive             | Authorization letter is not Embassy Attested.  |  |

|   |   |                    |   |   |   |   |   |   |   |   |   |   |   |   |    |   |   |   |   |   |   |   |   |                |  |  |
|---|---|--------------------|---|---|---|---|---|---|---|---|---|---|---|---|----|---|---|---|---|---|---|---|---|----------------|--|--|
|   | 7 | IBS Pharmaceutical | Autoclave Automatic 100L, Model:PS/QVA/SH100, Make:Priorclave UK.   | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | Non Responsive | Not according to specification, Manual Doors quoted.   |  |
|   | 8 | Advance Systems    | Autoclave Automatic 100L Model: 5075ELC G. Make: Tuttnauer, Netherland  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | Non Responsive | Quoted temperature, pressure, automatic doors not supported by broucher for the quoted model   |  |
| 4 | 1 | Medco              | Autoclave Automatic 150 L Model: BKQ-Z 200(H) Make Biobase China  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | Non Responsive | Not according to advertised specification, Pressure is less then Required. Quoted Manual doors |  |
|   | 2 | Clinical Life Inc  | Autoclave Automatic 150 Liters Model Name: Steri 21 Model number: 50.50.66 Manufacturer: A.J.Costa Origin: Portagal Ro system Model: DM-50-2A   | 2 | 2 | 5 | 0 | 5 | 5 | 5 | 5 | 0 | 2 | 2 | 30 | 0 | 0 | 4 | 0 | 0 | 0 | 0 | 4 | 71             | Responsive   | NIL  |
|   | 3 | Delta Plus Pvt Ltd | Autoclave Automatic 150 Liters Model: Azteca AC-575 Manufacturer: Celltron Medical Technology kft Origen,p Hungry (E.U) RO System: FIZICO 100L Manufacturer: FIZICO-Z.A Engineers Lahore (Dow USA parts)  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0              | Non Responsive   | Authorization letter is not Embassy Attested.  |
|   | 4 | RAY TECHNOLOGIES   | AUTOCLAVE AUTOMATIC 150 LITERS MODEL:HGS-133 HIRAYAMA MAKE:JAPAN  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0              | Non Responsive   | Not according to specifications. Separate Steam Generator Quoted                             |
|   | 5 | IBS Pharmaceutical | Autoclave Automatic 150L, Model:PS/QVA/SH150, Make:Priorclave UK.   | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0              | Non Responsive   | Not according to specification, Manual Doors quoted.   |
|   | 6 | Advance Systems    | Autoclave Automatic 150L Model: 5075ELC G. Make: Tuttnauer, Netherland  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0              | Non Responsive   | Quoted temperature, pressure, automatic doors not supported by broucher for the quoted model |
|   | 7 | KASBN              | AUTOCLAVE AUTOMATIC 150LTR MODEL: AS-449 WPA MAKE: SMSM POLAND  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0              | Non Responsive   | Authorization letter is not Embassy Attested.  |
| 5 | 1 | Medco              | Autoclave Automatic 200 L Model: BKQ-Z 300(H) Make Biobase China  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | Non Responsive | Not according to advertised specification, Pressure is less then Required. Quoted Manual doors |  |
|   | 2 | Clinical Life Inc  | Autoclave Automatic 200 Liters Model Name: Steri 21 Model number: 50.50.100 Manufacturer: A.J.Costa Origin: Portagal Ro system Model: DM-50-2A  | 2 | 2 | 5 | 0 | 5 | 5 | 5 | 5 | 0 | 2 | 2 | 30 | 0 | 0 | 4 | 0 | 0 | 0 | 0 | 4 | 71             | Responsive   | NIL  |
|   | 3 | Delta Plus Pvt Ltd | Autoclave Automatic 200 Liters Model: Azteca AC-5100 Manufacturer: Celltron Medical Technology kft Origen,p Hungry (E.U) RO System: FIZICO 100L Manufacturer: FIZICO-Z.A Engineers Lahore (Dow USA parts) | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0              | Non Responsive   | Authorization letter is not Embassy Attested.  |
|   | 4 | Total Technology   | Autoclave Automatic 200 Liters Model: Clinidav 45M Make: Melag Germany  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0              | Non Responsive   | Not according to specification, Pressure is less then required.                              |
|   | 5 | IBS Pharmaceutical | Autoclave Automatic 200L, Model:PS/QVA/SH200, Make:Priorclave UK.   | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0              | Non Responsive   | Not according to specification, Manual Doors quoted.   |
|   | 6 | Advance Systems    | Autoclave Automatic 200L Model: 50125GS-1A-380. Make: Tuttnauer, Netherland   | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0              | Non Responsive   | Quoted pressure, not supported by broucher for the quoted model                              |

|    |   |                               |   |   |   |   |   |   |   |   |   |   |   |   |   |    |    |   |   |   |   |   |   |   |    |            |            |                |  |
|----|---|-------------------------------|---|---|---|---|---|---|---|---|---|---|---|---|---|----|----|---|---|---|---|---|---|---|----|------------|------------|----------------|--|
| 6  | 1 | Vertex Medical (PVT) LTD      | Baby Suction Machine Model Basic Make Medla AG Switzerland  | 2 | 2 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 0 | 0 | 2  | 29 | 0 | 4 | 4 | 0 | 0 | 3 | 0 | 4  | 80         | Responsive | Nil            |  |
|    | 2 | SUDAIS ASSOCIATES             | BABY SUCTION MACHINE MODEL:MS-700 MAKE: M.TECHNOLOGY KOREA  | 2 | 2 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 0 | 2 | 2  | 30 | 0 | 0 | 4 | 0 | 0 | 0 | 0 | 4  | 76         | Responsive | NIL            |  |
|    | 3 | KASBN                         | BABY SUCTION MACHINE MODEL:FV SERIES SM1240 MAKE ORDISI SPAIN EUROPE  | 2 | 2 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 0 | 2 | 2  | 30 | 0 | 0 | 4 | 0 | 0 | 2 | 0 | 4  | 78         | Responsive | NIL            |  |
|    | 4 | IBS Pharmaceutical            | Baby Suction Machine, Model:III Negative Pressure, Make: ARI China.   | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0          | 0          | Non Responsive | Not according to specification Capacity and quantity of Suction Bottle is less then Required.                |
|    | 5 | Al- Kareem Medical Technology | Baby Suction Machine, Model:MEVACS M20, Make:Medist Slovakia  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0          | 0          | Non Responsive | Authorization letter is not Embassy Attested.  |
| 7  | 1 | Blazon Diagnostic             | BLOOD BAG CENTRIFUGE MODEL: DP-206SR 12 B MAKE :PRESVAC SRL Argentina   | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0          | 0          | Non Responsive | Authorization letter is not Embassy Attested.  |
|    | 2 | Medco                         | Blood Bag Centrifuge Model: BKC-886A Make: Biossese China   | 2 | 2 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 0 | 0 | 0 | 30 | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 4 | 68 | Responsive | Nil        |                |  |
|    | 3 | IBS Pharmaceutical            | Blood Bag Centrifuge, Model: DL-GLM, Make:ARI China.  | 2 | 2 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 0 | 2 | 2 | 30 | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 4 | 72 | Responsive | Nil        |                |  |
| 8  | 1 | Blazon Diagnostic             | BLOOD CELL COUNTER MODEL MODEL: CC-1431 CAT 7121431 MAKE: JP SELECTA SPAIN  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0          | 0          | Non Responsive | Authorization letter is not Embassy Attested.  |
| 9  | 1 | ENDO_KARE                     | ENT Examination Chair Model SIT 4 OTOPRONT GERMAN   | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0          | 0          | Non responsive | Authorization letter is not Embassy Attested.  |
|    | 2 | MEDIUNE TECHNOLOGY            | ENT EXAMINATION CHAIR MODEL:MC 4000A MAKE: MEDSTAR CO LIMITED KOREA   | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0          | 0          | Non Responsive | Authorization letter is not Embassy Attested.  |
|    | 4 | Radiant Medical Pvt Ltd       | ENT Examination Chair, Model: 45A, Make: Heinemann Germany.   | 2 | 2 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 0 | 5 | 0 | 2  | 29 | 0 | 0 | 4 | 0 | 0 | 0 | 0 | 0  | 64         | Responsive | Nil            |  |
| 10 | 1 | ENDO_KARE                     | ENT Examination unit with Cautery section Model SMART OTOPRONT GERMAN   | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0          | 0          | Non responsive | Authorization letter is not Embassy Attested.  |
|    | 3 | MEDIUNE TECHNOLOGY            | ENT Examination unit with Cautery section MODEL:UE 3000 PLUS MAKE: MEDSTAR CO LIMITED KOREA                         | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0          | 0          | Non Responsive | Authorization letter is not Embassy Attested.  |
|    | 2 | Radiant Medical Pvt Ltd       | ENT Examination unit with Cautery section. Model: Modula 4.5A Europa. Make: Heinemann Germany                       | 2 | 2 | 5 | 5 | 5 | 5 | 5 | 5 | 0 | 5 | 0 | 2 | 28 | 0  | 0 | 4 | 0 | 0 | 0 | 0 | 0 | 0  | 63         | Responsive | Nil            |  |
|    | 3 | BioCare                       | ENT Examination unit with Cautery section. Model: Otlaphari compact plus Dantschke Medizintechnik Germany           | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0          | 0          | Non responsive | Authorization letter is not Embassy Attested.  |
| 11 | 1 | Pakistan Marketing Services   | Glucometer Make Viva Chek Laboratories Inc China  | 2 | 2 | 0 | 5 | 5 | 5 | 5 | 5 | 5 | 0 | 0 | 2 | 30 | 0  | 4 | 4 | 0 | 0 | 2 | 2 | 4 | 77 | Responsive | Nil        |                |  |
|    | 2 | Qazzafi Surgical              | Glucometer Model: Easy Gluco Make: Infopia Korea  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0          | 0          | Non Responsive | Authorization letter is not Embassy Attested.  |
| 12 | 1 | IBS Pharmaceutical            | Gynae Vacuum Suction Apparatus/Vacuum Extractor With Trolley, Model:23C.III Mobile Electric Suction Make:ARI China. | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0          | 0          | Non Responsive | Not according to specification Capacity of Suction Bottle is less then Required. Suction currents not quoted |
|    | 2 | Al- Kareem Medical Technology | Gynae Vacuum Suction Apparatus/Vacuum Extractor With Trolley, Model:MEVACS M90, Make: Medist, Slovakia.             | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0          | 0          | Non Responsive | Authorization letter is not Embassy Attested.  |
|    | 3 | SUDAIS ASSOCIATES             | GYNAE VACCUME SUCTION WITH TROLLY MODEL: MS-700 MAKE: M.TECHNOLOGY KOREA  | 2 | 2 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 0 | 2 | 2 | 30 | 0  | 0 | 4 | 0 | 0 | 0 | 0 | 4 | 76 | Responsive | NIL        |                |  |
|    | 3 | KASBN                         | GYNAE VACCUME SUCTION MODEL:FV SERIES SM1240 MAKE ORDISI SPAIN EUROPE   | 2 | 2 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 0 | 2 | 2 | 29 | 0  | 0 | 4 | 0 | 0 | 2 | 0 | 4 | 77 | Responsive | NIL        |                |  |
|    | 4 | Vertex Medical (PVT) LTD      | Gynae Vacuum Suction Apparatus/ vacuum extractor with accessories Model: DOMINATE FLEX Medela AG Switzerland        | 2 | 2 | 5 | 5 | 5 | 5 | 5 | 5 | 5 | 0 | 0 | 2 | 29 | 0  | 4 | 4 | 0 | 0 | 3 | 0 | 4 | 80 | Responsive | Nil        |                |  |
| 13 | 1 | Al- Kareem Medical Technology | Humidifier Model: AIRnava, Make: WILMed, Germany.   | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0          | 0          | Non Responsive | Authorization letter is not Embassy Attested.  |
| 14 | 1 | SUDAIS ASSOCIATES             | INDIRECT OPHTHALMOSCOPE MODEL:OP-60 MAKE: RAY VISION CHINA  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0          | 0          | Non Responsive | Not according to specifications, Light intensity is not mentioned in catalogue                               |
|    | 1 | LATIF BROTHERS                | INDIRECT OPHTHALMOSCOPE MODEL:IO-ALPIA MAKE: NEITZ JAPAN  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0          | 0          | Non Responsive | Not according to specifications, Light intensity 2000 lux not mentioned                                      |
|    | 2 | Radiant Medical Pvt Ltd       | Indirect Ophthalmoscope, Model: 12500-D, Make: Welch Allyn Inc USA.   | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0          | 0          | Non Responsive | Authorization letter is not Embassy Attested.  |



**Technical Evaluation Criteria for the Purchase of  
Category E Equipment (Section-IV of Statement of Requirement) for the financial year 2017-18**

| S# | 1  | 2  | 3  |   |  |   |  |   |  |   |   | 4   |   |   |  |   |   |   |       |                | 22  | 23 |
|----|--|--|--|---|--|---|--|---|--|---|---|---|---|---|--|---|---|---|-------|----------------|---|----|
|    |  |  | Firm Evaluation Parameters   |   |  |   |  |   |  |   |   | Product Evaluation Parameter  |   |   |  |   |   |   |       |                |   |    |
|    |  |  | Personnel/<br>Human Resource   |   |  | Workshop facility<br>Testing/Calibration of Equipment |  |   | Manufacturer<br>Performance  |   |   | Conformanc<br>e To<br>Specification   |   |   | Requisite Certifications                               |   |   | Product Performance Index   |       |                |   |    |
|    |  |  | Audited Balance Sheets & Income Tax return for last three years (each year 2 marks with Sales above Rs.200.0 million per year) | Financial Soundness exhibiting last three years bank statement showing capability to execute Contract. (Rs.100 M/PKR or above) (03 marks per certificate) | Minimum 03 contracts of fabrication completed. (Min. Rs.200.00 M or above (02 points per contract. Copy of work order be provided) | 35 Facility at Peshawar- Khyber Pakhtunkhwa           | After Sales satisfactory certificate from at least 03 users of same items (Per certificate 2 Nos.) | Detail of Spare parts availability on firm's pad (for the quoted items) | Mobile After Sales Vans in Peshawar for onsite repair & maintenance of Vehicles & Equipment. | Valid ISO 14001 (Environment Certificate) | Valid ISO 9001 Quality Management Certificate | Fully compliance with the required specifications as per statement of Requirement (Up to a maximum of four Minor deviations may be accommodated subject to the condition that main function and performance in any aspect would not affect. However, up to four marks will be deducted) | Authorization Certificate of Manufacturer if the bidder is not a Manufacturer (i.e. Chassis / Superstructure). The manufacturer should meet all the criteria given in these documents.) | Registration with PEC in relevant category i.e. fabrication & Code i.e. MEO6 (Up to Project Value of 300 Million) | Valid Registration with Sales & Income Tax Authorities | Weightage for local Pakistani original manufacturer (OEM certificate from PEC or any other relevant body) | Minimum 05 years of successful fabrication. (Satisfactory performance certificate from at least 03 users. (2 marks per certificate) | Warranty Period three years both with spare parts and services & Next two year services without parts, from the date of handing/taking over | Total | Status         |   |    |
|    |  |  | 6  | 6   | 9  | 6   | 6  | 5   | 8  | 3   | 3   | 30  | No marks, being mandatory parameter   |   |  | 6   | 6   | 6   | 100   |                |   |    |
| 1  | Toyota Khyber Motors                       | Ambulances   | 0  | 0   | 0  | 0   | 0  | 0   | 0  | 0   | 0   | 0   | 0   | 0   | 0  | 0   | 0   | 0   | 0     | Non Responsive | Not registered with PEC in ME06 code. Catalog ambulatory items not attached. Also make/model of ambulatory items not available. |    |
|    | Colibrative Heavy Industries (Pvt) Limited | Ambulance Hino Dutro WU-640R, 4 x 2, Origin Japan Assembled in Pakistan.     | 6  | 6   | 9  | 6   | 6  | 5   | 8  | 3   | 3   | 30  | Y   | Y   | Y  | -   | 6   | 0   | 88    | Responsive     | Nil   |    |
|    | Khawaja Brothers                           | TOYOTA Hiace Dual A/C High Roof 2.5L   | 0  | 0   | 0  | 0   | 0  | 0   | 0  | 0   | 0   | 0   | 0   | 0   | 0  | 0   | 0   | 0   | 0     | Non responsive | Authorization not from Indus Motors (2) fabricator not register with PEC in the relevant code ME06                              |    |
|    | Colibrative Heavy Industries (Pvt) Limited | Ambulance Toyota Hiace High Roof (4x2) with fabrication and ambulatory items | 0  | 0   | 0  | 0   | 0  | 0   | 0  | 0   | 0   | 0   | 0   | 0   | 0  | 0   | 0   | 0   | 0     | Non responsive | Authorization not from Indus Motors   |    |
| 2  | Colibrative Heavy Industries (Pvt) Limited | Ambulance Toyota Hilux 4x4 single cabine with fabrication and                | 0  | 0   | 0  | 0   | 0  | 0   | 0  | 0   | 0   | 0   | 0   | 0   | 0  | 0   | 0   | 0   | 0     | Non responsive | Authorization not from Indus Motors   |    |
|    | Khawaja Brothers                           | TOYOTA Hilux 4x4 Single Cabin Pickup   | 0  | 0   | 0  | 0   | 0  | 0   | 0  | 0   | 0   | 0   | 0   | 0   | 0  | 0   | 0   | 0   | 0     | Non responsive | Authorization not from Indus Motors (2) fabricator not register with PEC in the relevant code ME06                              |    |



**Technical Evaluation Criteria for the Purchase of  
Category F Equipment (Left over items) (Section-IV of Statement of Requirement) for the financial year 2017-18**

| S#    | No of Firms submitted Bids | Firms submitted Bids   | 2<br>Name of equipment | Firm Evaluation Parameters  |   |       |  |   |   |    |  |                           | Product Evaluation Parameter  |   |   |   |  |                | 19<br>Status   | 20<br>Remarks  |  |  |  |  |  |  |
|-------|----------------------------|--|------------------------|---|---|-------|--|---|---|----|--|---------------------------|---|---|---|---|--|----------------|--|----------------|--|--|--|--|--|--|
|       |                            |  |                        | Personnel/<br>Human Resource  |   |       | Workshop<br>facility<br>Testing/Calib<br>ration of<br>Equipment                      |   | Manufac<br>turer<br>Performance               |    | Conformance To<br>Specification                |                           | Product Certification   |   |   | Product Local Performance   |  |                |  |                |  |  |  |  |  |  |
|       |                            |  |                        | Audited Balance Sheets - Income Tax return & Sales tax returns for last three years 03 marks for each year (Rs 1.00 Million or above) |   |       | Mobile after sales vans in Peshawar for on sight repair                              |   | Local Training of Technical Staff (02 person) |    | Availability of workshop in Khyber Pakhtunkhwa |                           | Availability of workshop at National level  |   | Valid ISO 14001 (Environment Certificate) |   | Valid ISO 9001 - 2008 Quality Management Certificate |                |  |                | Fully compliance with the required specifications as per statement of Requirement (Up to a maximum of four Minor deviations may be accommodated subject to the condition that main function and performance in any aspect would not affect. However, up to four marks will be deducted)  |  |  |  |  |  |
|       |                            |  |                        | Full load test report from manufacturer   |   |       | Authorization certificate from the manufacturer if the bidders is not a manufacturer |   |   |    |  |                           | Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) |   |   | In case of manufacturer provision of OEM certificate. OEM means that whole unit must be packaged by the manufacturer of the engine & no third party packaging is allowed. |  |                |  |                | Registration with PEC in relevant category & Codes (Compressor generator & MED3), Applicable in case of authorized agent   |  |  | Down time for response to the fault<br>02 working days from first communication from PE 10<br>03 working days from first communication from PE 06<br>05 working days from first communication from PE 02 |  |  |
| Total |                            |  | Total                  |   |   | Total |  |   | Total   |    |  | Total                     |   |   | Total                                     |   |  | Total          |  |                |  |  |  |  |  |  |
| 1     | IBS Pharmaceutical         | 40 KVA Generator. Model: XHX-C40 Make: Fujian Xinheng Xin Motors China   | 9                      | 7   | 4 | 8     | 6  | 3 | 3   | 40 | 6  | No marks, being mandatory |   |   | 10  | 4   | 100  | Non Responsive | (1) Registration with PEC in ME 03 code is not provided. (2) Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) not provided. |                |  |  |  |  |  |  |
| 2     | Khawaja Brothers           | Generator 40KVA, with Parkins UK Engine and MECC-ALTE Alternator Genset Model ERP-40(PERKINS 1103A-33TG1) Make Erga Genset Spain | 0                      | 0   | 0 | 0     | 0  | 0 | 0   | 0  | 0  | 0                         | 0   | 0 | 0   | 0   | 0  | 0              | 0  | Non responsive | Authorization not embassy attested   |  |  |  |  |  |
| 3     | MIA Energy (PVT) LTD.      | 40 KVA Generator. Model: KGP45.  | 0                      | 0   | 0 | 0     | 0  | 0 | 0   | 0  | 0  | 0                         | 0   | 0 | 0   | 0   | 0  | 0              | 0  | Non Responsive | (1) Authorization not Embassy Attested. (2) Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) not provided   |  |  |  |  |  |
| 4     | Millat Tractors Limited    | 40KVA Generator  | 0                      | 0   | 0 | 0     | 0  | 0 | 0   | 0  | 0  | 0                         | 0   | 0 | 0   | 0   | 0  | 0              | 0  | Non Responsive | Mandatory documents i.e (1). Technical offer containing details i.e construction of foundation pad, ATS Panel, weather and sound proof canopy, installation and testing of the unit, earthing network, gauge and description of cable not available. (2) No Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) available. |  |  |  |  |  |
| 5     | Orient Energy              | 40KVA Diesel Generator. Model: C44D5. Make: Cummins, UAE   | 9                      | 0   | 4 | 8     | 6  | 2 | 2   | 40 | 0  | Yes                       | Yes   | - | Yes                                       | 10  | 4  | 85             | Responsive   | Nil            |  |  |  |  |  |  |

|   |                            |   |   |   |   |   |   |   |   |    |   |     |     |   |     |    |   |    |            |                |   |
|---|----------------------------|---|---|---|---|---|---|---|---|----|---|-----|-----|---|-----|----|---|----|------------|----------------|---|
| 6 | POWER ZONE                 | GENERATOR 45 kVA<br>MODEL: -PZ 50 MAKE : NOT MENTIONED  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0   | 0   | 0 | 0   | 0  | 0 | 0  | 0          | Non Responsive | 1.Authorization for engine is from third party that is Mayphil in stead of Perkins. 2.Authorization for alternator is not available. 3. No Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical ( Soft & Hard Copy) available.  |
| 7 | Sherazi Trading Co Pvt Ltd | 40KVA Engine Model:<br>Cummins 14BT3.9G2<br>Alternator: Stamford Control<br>Module: DSE-6020 Powered<br>by AKSA Turkey From China<br>Plant        | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0   | 0   | 0 | 0   | 0  | 0 | 0  | 0          | Non Responsive | Mandatory documents, 1. Embassy Attested authorization is not available. 2. No Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical ( Soft & Hard Copy) available.  |
| 1 | IBS Pharmaceutical         | 60 KVA Generator. Model:<br>XHX-C60 Make: Fujian<br>Xinheng Xin Motors China  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0   | 0   | 0 | 0   | 0  | 0 | 0  | 0          | Non Responsive | (1) Registration with PEC in ME 03 code is not provided. (2) Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical ( Soft & Hard Copy) not provided.   |
| 2 | Khawaja Brothers           | Generator 60KVA, with<br>Parkins UK Engine and MECC-<br>ALTE Alternator Genset<br>Model ERP-60S(PERKINS<br>1103A-33TG3) Make Erga<br>Genset Spain | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0   | 0   | 0 | 0   | 0  | 0 | 0  | 0          | Non responsive | Authorization not embassy attested  |
| 3 | MIA Energy (PVT) LTD.      | 60 KVA Generator. Model<br>KGP60.   | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0   | 0   | 0 | 0   | 0  | 0 | 0  | 0          | Non Responsive | (1) Authorization not Embassy Attested.<br>(2) Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical ( Soft & Hard Copy) not provided  |
| 4 | Millat Tractors Limited    | 60KVA Generator Model:<br>1103A-33TG2 (Perkins)<br>Alternator: Leroy Somer LSA<br>42.3  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0   | 0   | 0 | 0   | 0  | 0 | 0  | 0          | Non Responsive | Mandatory documents i e (1). Technical offer containing details i.e construction of foundation pad, ATS Panel, weather and sound proof canopy, installation and testing of the unit, earthing network, gauge and description of cable not available. 2. Embassy Attested authorization is not available. 3. ME03 not available. 4. No Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical ( Soft & Hard Copy) available. |
| 5 | Orient Energy              | 60KVA Diesel Generator<br>Model:C66DS.<br>Make:Cummins, UAE   | 9 | 0 | 4 | 8 | 6 | 2 | 2 | 40 | 0 | Yes | Yes | - | Yes | 10 | 4 | 85 | Responsive | Null           |   |
| 6 | POWER ZONE                 | GENERATOR 60 kVA<br>MODEL: -PZ 65 MAKE : NOT MENTIONED  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0   | 0   | 0 | 0   | 0  | 0 | 0  | 0          | Non Responsive | 1.Authorization for engine is from third party that is Mayphil in stead of Perkins. 2.Authorization for alternator is not available. 3. No Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical ( Soft & Hard Copy) available.  |
| 7 | Sherazi Trading Co Pvt Ltd | 60KVA Engine Model:<br>Cummins 4BT3.9G2<br>Alternator: Stamford Control<br>Module: Deepsea module<br>Powered by AKSA Turkey From<br>China Plant   | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0   | 0   | 0 | 0   | 0  | 0 | 0  | 0          | Non Responsive | Mandatory documents, 1. Embassy Attested authorization is not available. 2. No Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical ( Soft & Hard Copy) available.  |

|   |   |                            |  |   |   |   |   |   |   |   |    |   |     |     |   |     |    |   |                |   |  |
|---|---|----------------------------|--|---|---|---|---|---|---|---|----|---|-----|-----|---|-----|----|---|----------------|---|--|
| 3 | 1 | MIA Energy (PVT) LTD.      | 100 KVA Generator. Model KGP100.   | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0   | 0   | 0 | 0   | 0  | 0 | 0              | Non Responsive  | (1) Authorization not Embassy Attested.<br>(2) Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical ( Soft & Hard Copy) not provided   |
|   | 2 | IBS Pharmaceutical         | 100 KVA Generator. Model: XHX-C100 Make: Fujian Xinheng Xin Motors China   | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0   | 0   | 0 | 0   | 0  | 0 | 0              | Non Responsive  | (1) Registration with PEC in ME 03 code is not provided. (2) Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical ( Soft & Hard Copy) not provided.  |
|   | 3 | Khawaja Brothers           | Generator 100KVA with Parkins UK Engine and MECC-ALTE Alternator Genset Model ERP-100(PERKINS 1104C-44TAG2) Make Erga Genset Spain               | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0   | 0   | 0 | 0   | 0  | 0 | 0              | Non responsive  | Authorization not embassy attested   |
|   | 4 | Millat Tractors Limited    | 100KVA Generator Model: 1104A-44TG2 (Perkins) Alternator: Leroy Somer TAL040-TAL042-TAL044   | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0   | 0   | 0 | 0   | 0  | 0 | 0              | Non Responsive  | Mandatory documents i.e (1). Technical offer containing details i.e construction of foundation pad, ATS Panel, weather and sound proof canopy, installation and testing of the unit, earthing network, gauge and description of cable not available. 2. Embassy Attested |
|   | 5 | Orient Energy              | 100KVA Diesel Generator Model: C110D5. Make: Cummins, China.   | 9 | 0 | 4 | 8 | 6 | 2 | 0 | 40 | 6 | Yes | Yes | - | Yes | 10 | 4 | 89             | Responsive  | Null   |
|   | 6 | POWER ZONE                 | GENERATOR 100 kVA MODEL: -PZ 100 MAKE : NOT MENTIONED  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0   | 0   | 0 | 0   | 0  | 0 | 0              | Non Responsive  | 1.Authorization for engine is from third party that is Mayphil instead of Perkins. 2.Authorization for alternator is not available. 3. No Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical ( Soft & Hard Copy) available.        |
|   | 7 | Sherazi Trading Co Pvt Ltd | 100KVA Engine Model: Cummins 6BT5.9G2 Alternator: Stamford Model: UCI274C Control Module: Deepsea module Powered by AKSA Turkey From China Plant | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0   | 0   | 0 | 0   | 0  | 0 | 0              | Non Responsive  | Mandatory documents, 1. Embassy Attested authorization is not available. 2. No Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical ( Soft & Hard Copy) available.   |
|   | 1 | IBS Pharmaceutical         | 200 KVA Generator. Model: XHX-C200 Make: Fujian Xinheng Xin Motors China   | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0   | 0   | 0 | 0   | 0  | 0 | Non Responsive | (1) Registration with PEC in ME 03 code is not provided. (2) Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical ( Soft & Hard Copy) not provided. |  |
|   | 2 | Khawaja Brothers           | Generator 200KVA with Parkins UK Engine and MECC-ALTE Alternator Genset Model ERP-200(PERKINS 1106A-70TAG4) Make Erga Genset Spain               | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0   | 0   | 0 | 0   | 0  | 0 | Non responsive | Authorization not embassy attested  |  |
|   | 3 | MIA Energy (PVT) LTD.      | 200 KVA Generator. Model KGP200.   | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0   | 0   | 0 | 0   | 0  | 0 | Non Responsive | (1) Authorization not Embassy Attested.<br>(2) Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical ( Soft & Hard Copy) not provided                |  |

|   |                            |  |   |   |   |   |   |   |   |    |   |     |     |   |     |    |   |    |            |      |                |   |
|---|----------------------------|--|---|---|---|---|---|---|---|----|---|-----|-----|---|-----|----|---|----|------------|------|----------------|---|
| 4 | Millat Tractors Limited    | 200KVA Generator Model:<br>1106A-70TG4 (Perkins)<br>Alternator: Leroy Somer<br>TAL046-TAL047-TAL049  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0   | 0   | 0 | 0   | 0  | 0 | 0  | 0          | 0    | Non Responsive | Mandatory documents i.e (1). Technical offer containing details i.e. construction of foundation pad, ATS Panel, weather and sound proof canopy, installation and testing of the unit, earthing network, gauge and description of cable not available. 2. Embassy Attested authorization is not available. 3. ME03 not available. 4. No Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) available. |
| 5 | Orient Energy              | 200KVA Diesel Generator<br>Model: C220D5.<br>Make:Cummins, China   | 9 | 0 | 4 | 8 | 6 | 2 | 0 | 40 | 6 | Yes | Yes | - | Yes | 10 | 4 | 89 | Responsive | Null |                |   |
| 6 | POWER ZONE                 | GENERATOR 200 kVA<br>MODEL: -PZ 200 MAKE :<br>NOT MENTIONED  | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0   | 0   | 0 | 0   | 0  | 0 | 0  | 0          | 0    | Non Responsive | 1.Authorization for engine is from third party that is Mayphil instead of Perkins. 2.Authorization for alternator is not available. 3. No Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) available.  |
| 7 | Sherazi Trading Co Pvt Ltd | 200KVA Engine Model:<br>Cummins 8CTAA8.9G2<br>Alternator: Stamford Control<br>Module: Deepsea module<br>Powered by AKSA Turkey From<br>China Plant | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0  | 0 | 0   | 0   | 0 | 0   | 0  | 0 | 0  | 0          | 0    | Non Responsive | Mandatory documents, 1. Embassy Attested authorization is not available. 2. No Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) available.   |